



Calibration Report



Test # 164595

United States - Main Office

3839 County Road 116
Medina, MN 55340 USA

www.intercompco.com

TEL. (763) 476-2531 FAX (763) 476-2613

United Kingdom

The Old Barn, Manor Farm
Manor Rd., Shurlock Row

Berkshire RG10 0PY

TEL. 44-118-932-0578 FAX 44-118-932-1034

Tester	KR			Model #	SW500		
Date	06/May/2021			Scale/Cell ID #			
Indicator				Capacity	1500 Lbs.		
Serial#	0123MA13032			Graduation (d)	1		
Part #	170125			Re-Cal Date	06/May/2022		
Temperature	18.2 °C	Humidity	48.7 %	Ack #		PO #	

Accuracy: +/- 0.1 % of applied load or +/- 1 Lbs. - whichever is greater

As Left: In Tolerance

Weight(Lbs.)	Left Front(Lbs.)	Left Rear(Lbs.)	Right Front(Lbs.)	Right Rear(Lbs.)
0	0	0	0	0
200	200	200	200	200
500	500	500	500	500
1000	999	1000	999	999
0	0	0	0	0

Intercomp Company does hereby certify the above listed instrument meets or exceed all published Specifications and has been calibrated using standards whose accuracies are directly traceable to the U.S. National Institute of Standards and Technology.

This document is not to be reproduced except in full without the written permission of Intercomp

Calibration Procedure:

Model #	Test/Trace #	Re-Cal Date
CS10-2K-B000	Tovey 50244-1	15/Mar/2022

Tested By: 

Title: Service Technician

Approve By: 

Title: Director